

Case Studies: DTSC's Environmental Chemistry Lab's (ECL) testing of two types of consumer products

Mui Koltunov, Ph.D.

Nan Xie, Ph.D.

Wednesday, June 26, 2025

1,4-Dioxane in Personal Care Products



BD02813



BD02814



BD02815



BD02816



BD02817



BD02818



BD02819



BD02820



BD02821



BD02822



BD02823



BD02824



BD02825



BD02826



BD02827



BD02828



BD02829



BD02830



BD02831



BD02832



BD02833



BD02834






BD02835



BD02836

One Possible Technique – Sorbent Pens




SPME-GC/MS/MS Technique



- Solvent-free
- Non-exhaustive
- Equilibrium technique
- Minimal sample-prep
- High-throughput

SPME-GC/MS/MS Technique





Instrument Checks Acceptance Criteria

Instrument Measure	Frequency	Requirement
Calibration	As needed	<ul style="list-style-type: none">Avg. Response Factor %RSD $\leq 20\%$Accuracy $\pm 30\%$ of True Value, except LOQ where Accuracy $\pm 50\%$ of True Value
ICV	Immediately after calibration	<ul style="list-style-type: none">ICV $\pm 30\%$ True Value
CCV	Before analysis and after every 10 analytical runs (excluding any blanks).	<ul style="list-style-type: none">CCV $\pm 20\%$ True Value
Internal Standard (ISTD)	Added to every sample, QC, calibration level and instrument check	<ul style="list-style-type: none">N/A
Retention Time (RT)	Evaluate in every sample	<ul style="list-style-type: none">ISTD RT ± 0.33 min to mid-point ICALAnalyte RT < 10 sec. to mid-point ICAL or first CCV

Quality Control Acceptance Criteria

Batch QC	Frequency	Requirement
MB	With every batch of 10 or fewer samples	<ul style="list-style-type: none"> Analyte \leq LOQ
LCS/LCSD	With every batch of 10 or fewer samples	<ul style="list-style-type: none"> % Recovery: refer to Control Limits table RPD of LCS/LCSD \leq 20%
MS/MSD	With every batch of 10 or fewer samples	<ul style="list-style-type: none"> Spike Recovery \pm 30% RPD of MS/MSD \leq 20%
Sample Duplicate	With every sample	<ul style="list-style-type: none"> RPD of Sample/Duplicate \leq 20%
LOQ	Per Customer Request	<ul style="list-style-type: none"> Spike Recovery \pm 50% RSD \leq 20%
QC/LCS	4 replicates annually for analyst not performing PT	<ul style="list-style-type: none"> % Recovery: refer to Control Limits table RPD \leq 20%
MU	At least 9 LCS for method verification / validation	<ul style="list-style-type: none"> Spike Recovery \pm 30%
PT	Per Customer Request	<ul style="list-style-type: none"> Within acceptable PT study range

1,4-dioxane in DTSC-tested products



Application Notes Available

1,4-dioxane in consumer products by SPME
and GC-QQQ



Analysis of 1,4-Dioxane in
Consumer Products by Solid Phase
Microextraction and Triple Quadrupole
GC/MS

1,4-dioxane in consumer products by
Headspace-Gas and GC-MS



Analysis of 1,4-Dioxane in Consumer
Products by Headspace-Gas
Chromatography/Mass Spectrometry

Total Fluorine in Carpets/Rugs by CIC


Fiber

- No known sources of inorganic fluorine (F).
- Assume total F = PFASs




Backing

- Strong possibility of inorganic F. e.g., limestone filler.
- May also contain organic F
 - PFASs from recycled content
 - PFASs from fiber not fully removed
 - PFASs used as extrusion agents during manufacturing
- Cannot assume that total F = PFASs



CIC Flowpath – Direct Inject for F Analysis



Instrument Checks Acceptance Criteria - CIC

Instrument Measure	Frequency	Requirement
Calibration	As needed	<ul style="list-style-type: none">• $R^2 \geq 0.99$• $\pm 10\%$ of True Value
ICV	Immediately after calibration	<ul style="list-style-type: none">• $\pm 30\%$ of True Value
CCV	Before sample analysis and at the end of analysis	<ul style="list-style-type: none">• $\pm 20\%$ of True Value
Check Compound	Before sample analysis and at the end of analysis	<ul style="list-style-type: none">• $\pm 30\%$ of True Value
CCB	Before sample analysis and at the end of analysis	<ul style="list-style-type: none">• $CCB < RL$
Retention Time (RT)	Evaluate in every sample	<ul style="list-style-type: none">• Within 15% of the mid-point ICAL or the CCV

Quality Control Acceptance Criteria - CIC

Batch QC	Frequency	Requirement
MB	Before sample analysis and at the end of analysis	<ul style="list-style-type: none"> Analyte < LOQ
LCS/LCSD	With every batch of 10 or fewer samples	<ul style="list-style-type: none"> % Recovery \pm 30% RPD of LCS/LCSD \leq 20%
MS/MSD	With every batch of 10 or fewer samples	<ul style="list-style-type: none"> Spike Recovery \pm 30% RPD of MS/MSD \leq 20%
Sample Duplicate	Each sample	<ul style="list-style-type: none"> RPD of Sample/Duplicate \leq 20%
LOQ	7 replicate samples annually	<ul style="list-style-type: none"> % Recovery \pm 50% RSD \leq 20%
QC/LCS	4 replicates annually	<ul style="list-style-type: none"> % Recovery \pm 30% RSD \leq 20%
MU	20 LCS for method verification / validation	<ul style="list-style-type: none"> N/A
MDL	Annually	<ul style="list-style-type: none"> N/A

Reprocessing table

	Determination start	Method	Ident	Sample type	Injecti...	Volume	Dil...	Sa...	Integ...	In...	Inf...	Info 3 I...	Value 1	Valu
1	2022-12-16 11:35:43 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
2	2022-12-16 12:17:49 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
3	2022-12-16 12:41:52 UTC-8	Full System Calibration	CCV1	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
4	2022-12-16 13:05:54 UTC-8	Full System Calibration	CCB1	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
5	2022-12-16 13:29:56 UTC-8	Full System Calibration	MB1	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
6	2022-12-16 13:53:58 UTC-8	Full System Calibration	LCS	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
7	2022-12-16 14:17:59 UTC-8	Full System Calibration	LCSD	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
8	2022-12-16 14:42:04 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
9	2022-12-16 15:06:30 UTC-8	Full System Calibration	BGA00730-A	Sample	1/1	250	1	1	Auto...	D...	Sp...		14.6	1
10	2022-12-16 15:30:56 UTC-8	Full System Calibration	BGA00730-A Dup	Sample	1/1	250	1	1	Auto...	D...	Sp...		17.6	1
11	2022-12-16 15:54:59 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
12	2022-12-16 16:19:22 UTC-8	Full System Calibration	BGA00731-A	Sample	1/1	250	1	1	Auto...	D...	Sp...		23	1
13	2022-12-16 16:43:40 UTC-8	Full System Calibration	BGA00731-A Dup	Sample	1/1	250	1	1	Auto...	D...	Sp...		18.6	1
14	2022-12-16 17:07:42 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
15	2022-12-16 17:32:05 UTC-8	Full System Calibration	BGA00732-A	Sample	1/1	250	1	1	Auto...	D...	Sp...		12.8	1
16	2022-12-16 17:56:31 UTC-8	Full System Calibration	BGA00732-A Dup	Sample	1/1	250	1	1	Auto...	D...	Sp...		17.8	1
17	2022-12-16 18:20:37 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
18	2022-12-16 18:44:54 UTC-8	Full System Calibration	BGA00733-A	Sample	1/1	250	1	1	Auto...	D...	Sp...		22.3	1
19	2022-12-16 19:09:19 UTC-8	Full System Calibration	BGA00733-A Dup	Sample	1/1	250	1	1	Auto...	D...	Sp...		20.2	1
20	2022-12-16 19:33:24 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1
21	2022-12-16 19:58:02 UTC-8	Full System Calibration	BGA00734-A	Sample	1/1	250	1	1	Auto...	D...	Sp...		27	1
22	2022-12-16 20:22:10 UTC-8	Full System Calibration	BGA00734-A Dup	Sample	1/1	250	1	1	Auto...	D...	Sp...		23.5	1
23	2022-12-16 20:46:18 UTC-8	Full System Calibration	blank	Sample	1/1	250	1	1	Auto...	D...	Sp...		200	1

Edit4

Results

Results

Anions

Component name	Retention time [min]	Height [µS/cm]	Area [(µS/cm) x min]	Concentration [ppm]	Percentual standard error [%]	Percentage standard deviation [%]	Spik
Fluoride	4.37	30.109	5.321	46.122	3.283	1.181	

Single results

1_Concentration Fluoride: 341.642 ppm

Accuracy: NA %

CV Volume water inlet: 1.027 mL

CV combustion time: 10.6 min

CV nested combustion: 1

CV total volume: 8.001 mL

CV volume initial absorber solution: 2.000 mL

CV volume rinsing after combustion: 2.921 mL

CV volume transfer absorber solution: 2.053 mL

Combustion time: invalid min

Nested combustion: 1

Nested combustion time: 10.0 min

Nested volume initial absorber solution: 2.000 mL

Nested volume rinsing after combustion: 3.108 mL

Nested volume transfer during combustion: 1.928 mL

Nested volume water inlet: 0.964 mL

Total volume: 0.000 mL

Evaluation parameters - Method from determination 21

Integration

Concentration unit ppm


Standards

Check standards


Spiking solutions

	Name	Standard 1	Standard 2	Standard 5	Standard 10	Standard 25	Standard 50	*
1	Fluoride	2	5	10	50	100	200	


Edit4




Total F in Carpet & Rug Fibers: Historical vs Compliance Samples




Total F in Carpet & Rug Fibers: Historical vs Compliance Samples (Positive Detections Only)




Total F in Carpet & Rug Backing: Historical vs. Compliance Samples



Total F in Carpet & Rug Backing: Historical vs Compliance Samples (Positive Detections Only)



Reporting



Reporting



- Case Narrative
- Acronyms
- Data Tables
- Preliminary Analysis
- Request Review
- Authorization Request
- Sample Analysis Request
- Sample Chain of Custody
- Sample Receipt Checklist
- Photos
- Airbill

Level 2 Report
(Sent to Requestor)

Everything in the Level 2 Report PLUS:

- Customer Technical Assistance Form
- Solution Preparation Forms
- Reference Material Certificates of Analysis
- Data Reduction Tables
- Pipette IDs used
- Chromatograms

Level 4 Report
(Stored by ECL)

Report Examples

- Level II Report – 1,4-dioxane in personal care product
- Level IV Report – 1,4-dioxane in personal care product

- Level II Report – Total fluorine in carpets and rugs
- Level IV Report – Total fluorine in carpets and rugs